

SAFETY RECALL

Certain Fluke 333, 334, 335, 336 and 337 Digital Clamp Meters

Manufactured between January 29, 2008 and February 27, 2009



Recalled clamp meters		
Model Family	Starting Serial No.	Ending Serial No.
Fluke-337 (All 337 models, including 337A)	96070000	98890607
Fluke-336 (All 336 models, including 336A)	96220000	98900400
Fluke-335 (All 335 models, including 335A)	97010000	98860104
Fluke-334 (All 334 models, including 334A)	97010000	98870101
Fluke-333 (All 333 models, including 333A)	97010000	98880240

Note: Serial numbers that begin with "S" or are followed by "R" are NOT included in the recall.

Potential short circuit in clamp meter may lead to inaccurate or no voltage reading on circuit energized with hazardous voltage.

Customers should stop using these clamp meters immediately and contact Fluke for instructions to return these models and receive a free replacement unit. Contact Fluke at:

- www.fluke.com/33Xrecall
- 1-888-98-FLUKE (35853), 7 a.m. to 4 p.m. Pacific Time Monday–Friday